

CERTIFICATE

Alternative Bowie & Dick Test System

BAG-AutoCheck-Kit

from

BAG Health Care GmbH, D-35423 Lich

It is herewith certified that the single use Bowie & Dick Test System BAG-AutoCheck-Kit (REF 7521) fulfils performance requirements after exposure to dry saturated steam when tested under the same conditions as described in European Norm EN 867-4 for the Standard Test Pack.

This is also to certify that above testing was conducted in a Steam Sterilizer according EN 285 and using Air Injection Apparatus.

BAG Health Care GmbH

August 2007



Dr. Lenard Mueller
Product Manager
Hygiene Monitoring