

CERTIFICATE

Cross-Checks F

It is herewith certified that the chemical indicator Cross-Checks F (REF 7533), manufactured by Steritec Products Mfg., Inc. and distributed by BAG Health Care GmbH fulfils the requirements of a **Class 3** indicator according to the International Norm **ISO 11140-1**.

The certification is limited to the use in steam/formaldehyde sterilization processes.

The product presents no health or physical hazards under normal conditions of use.

BAG Health Care GmbH

December 2007



Dr. Lenard Mueller
Product Manager
Hygiene Monitoring